



Slick Silsm LSR

Regulatory Testing

FDA Indirect food contact (21CFR177.2600)

Extractives Passed 10/2004

Composition Compliant

Acute Dermal Irritation

Passed 11/2004

USP Class VI

Passed 11/2004

Please note:

Compliance testing conducted with LSR Top Coat on LIM6040-D1 substrate cured in to ASTM "slabs."

No post baking on substrate and top coat.

5170 N. Northwest Highway
Chicago, IL • 60630
P • 773.427.2084 - F • 773.282.9155

www.surfacesolutionsgroup.com